



95069-002 95DLG-6A

-0- M TRTR

Instrument Conditions: HP5890-VG7035 GC/MS S.D.M., m/z 372.3756>217.1956

Method: DEFAULT, Low Scale=0.002 mV High Scale=1.321 mV

95069602, Analyzed: 05/01/98 by U.S.G.S. Box 25046 MS 977 Denver, CO 80225